



EMS System for Metropolitan Oklahoma City and Tulsa 2022 Medical Control Board Treatment Protocols



 **EMS SECTION**

Approved 5/11/22, Effective 5/11/22, replaces all prior versions

3K – NON-INVASIVE POSITIVE PRESSURE VENTILATION (NIPPV) ADULT & INTER- FACILITY PEDIATRIC

EMT

EMT-INTERMEDIATE 85

ADVANCED EMT

PARAMEDIC

Indications:

1. Dyspnea – Uncertain Etiology – Adult.
2. Dyspnea – Asthma – Adult.
3. Dyspnea – Chronic Obstructive Pulmonary Disease (COPD) – Adult.
4. Dyspnea – Congestive Heart Failure (CHF) – Adult.
5. Acute Allergic Reactions – Adult (Dyspnea).
6. Water Submersion Event – Adult (Dyspnea).
7. Pediatric Dyspnea - Inter-Facility Continuation of Care.

Contraindications:

1. Apnea.
2. Pediatric dyspnea- Non-Inter-Facility/Non-Continuation of Care.
3. Adult dyspnea of lesser severity able to be managed without NIPPV.
4. Adult dyspnea of greater severity requiring invasive airway management.
5. Altered mental status preventing patient cooperation with NIPPV.
6. Active or suspected impending emesis.
7. High risk of aspiration/Impaired gag reflex.
8. Facial trauma/features impairing a tight NIPPV mask-face seal.




EMS System for Metropolitan Oklahoma City and Tulsa 2022 Medical Control Board Treatment Protocols



Approved 5/11/22, Effective 5/11/22, replaces all prior versions

PROTOCOL 3K: Non – Invasive Positive Pressure Ventilation (NIPPV) - Adult & Inter-Facility Pediatric, cont.

Bi-Level/CPAP Ventilation Algorithm



Special Considerations/Complications

- > Patients requiring bronchodilator therapy?
 - ✓ Bronchodilators via nebulizer t-piece in line with NIPPV
- > It is very important to achieve a tight seal between face and NIPPV mask to deliver anticipated levels of NIPPV
- > Monitor closely for nausea/impending emesis – be prepared to quickly remove facemask to avoid aspiration of emesis



EMS System for Metropolitan Oklahoma City and Tulsa 2022 Medical Control Board Treatment Protocols



Approved 5/11/22, Effective 5/11/22, replaces all prior versions

PROTOCOL 3K: Non – Invasive Positive Pressure Ventilation (NIPPV) - Adult & Inter-Facility Pediatric, cont.

Technique (ZoLL Z Vent):

Circuits:

Infant/ Pediatric Circuit

Pediatric/ Adult Circuit

1. Zoll Z ventilator circuits feature a low dead space design that minimizes CO2 re-breathing.
2. Note: dead space (circuit and HME) should never be greater than **25%** of the patient's tidal volume (set or spontaneous).
3. The 2 standard ventilator circuits cover the range of patient from infant through adult.

➤ Pediatric/adult – patients 20 kg through adult, minimum tidal volume 200mL.

*****Ventilator use in pediatrics restricted to inter-facility transport only.*****

➤ Infant/pediatric – 5 through 30 kg, maximum tidal volume 300 mL.


*****Ventilator use in pediatrics restricted to inter-facility transport only.*****

halation
ive Line
Clear)

ine

Connections - check the ventilator for proper operation before connecting to patient:

Step 1: Connect ventilator circuit (use test lung whenever possible) oxygen hose to 55 psi regulated output.





EMS System for Metropolitan Oklahoma City and Tulsa 2022 Medical Control Board Treatment Protocols



Approved 5/11/22, Effective 5/11/22, replaces all prior versions

PROTOCOL 3K: Non - Invasive Positive Pressure Ventilation (NIPPV) - Adult & Inter-Facility Pediatric, cont.

Step 2: Power

a Self-Check and the internal

gins operation ult settings. performed every eafter or when an perature change

gs may be g operation at



- Vt: 450 ml
- BPM: 12
- I:E 1:3
- Mode: AC (V)

21%
35 cm H2O
5 cm H2O

Primary Parameter:



ed

diac

M